Rat Leukotriene D4 (LT-D4) ELISA Kit

Catalog No: #EK10008

SAB Signalway Antibody

Package Size: #EK10008-1 48T #EK10008-2 96T

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

_			4.5	
	esc	rın	TIC	۱'n
-	しつし	IIV	uv	и

Product Name	Rat Leukotriene D4 (LT-D4) ELISA Kit	
Brief Description	ELISA Kit	
Applications	ELISA	
Species Reactivity	Rat (Rattus norvegicus)	
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%	
	within the expiration date under appropriate storage condition.	
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,	
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China	
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage	
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).	

Application Details

Detect Range:7.8-500 pg/mL
Sensitivity:1.95 pg/mL
Sample Type:Serum, Plasma, Other biological fluids
Sample Volume: 1-200 μL
Assay Time:1-4.5h
Detection wavelength:450 nm

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate LT-D4 in samples. An antibody specific for LT-D4 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyLT-D4 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for LT-D4 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of LT-D4 bound in the initial step. The color development is stopped and the intensity of the color is measured. Product Overview: This assay employs the quantitative sandwich enzyme immunoassay technique. Antibody specific for LTD4 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any LTD4 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for LTD4 is added to the wells. After washing, avidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of LTD4 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only